

Notice of Allowability

Application No.

10/681,473

Examiner

Annette R. Reimers

Applicant(s)

MOUMENE ET AL.

Art Unit

3733

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to correspondence filed on August 01, 2006.
2. ☒ The allowed claim(s) is/are 1,2,7-13 and 16-20.
3. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☐ All b) ☐ Some* c) ☐ None of the:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

* Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.

THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
 5. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
 - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

1. ☐ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☐ Information Disclosure Statements (PTO/SB/08),
Paper No./Mail Date _____
4. ☐ Examiner's Comment Regarding Requirement for Deposit
of Biological Material
5. ☐ Notice of Informal Patent Application
6. ☐ Interview Summary (PTO-413),
Paper No./Mail Date _____
7. ☐ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other _____


EDUARDO C. ROBERT
SUPERVISORY PATENT EXAMINER

REASONS FOR ALLOWANCE

The following is an examiner's statement of reasons for allowance:

The claims in the instant application have not been rejected using prior art because no references, or reasonable combination thereof, could be found which disclose, or suggest a motion disc comprising: a) a first prosthetic vertebral endplate comprising: i) an first outer surface having a porous coating thereon and adapted to mate with a first vertebral body, ii) an first inner surface comprising an first peripheral surface and a first articulation surface, iii) a body portion connecting the first inner and outer surfaces, b) a second prosthetic vertebral endplate comprising: i) an second outer surface having a porous coating thereon and adapted to mate with a second vertebral body, and ii) an second inner surface comprising a second peripheral surface and a second articulation surface, iii) a body portion connecting the second inner and outer surfaces, c) an articulating core member adapted to maintain intervertebral spacing and comprising: i) a first articulation surface adapted for articulation with the first articulation surface of the first endplate, and ii) a second articulation surface adapted for articulation with the first articulation surface of the second endplate, wherein the articulating core member is disposed between the prosthetic endplates and oriented to produce a first articulation interface between the first articulation surface of the first endplate and the first articulation surface of the core member, and a second articulation interface between the first articulation surface of the second endplate and the second articulation surface of the core member, and d) a peripheral shock-absorbing component comprising: i) a first surface contacting the first peripheral surface of the first

Art Unit: 3733

endplate, and ii) a second surface contacting the second peripheral surface of the second endplate, wherein the core component is selected from the group consisting of polyethylene, PEEK, ceramic and metals, and wherein the shock-absorbing component comprises an elastomer, and wherein each peripheral surface of the shock-absorbing component is tenaciously attached to the respective inner surface of each endplate, as set forth in claim 1.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Annette R. Reimers whose telephone number is (571) 272-7135. The examiner can normally be reached on Monday-Friday.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached on (571) 272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3733

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

AR

AK


EDUARDO C. ROBERT
SUPERVISORY PATENT EXAMINER